

Fluke Thermography

Certificate of Calibration

Serial Number: **Ti9-11060227**

Model: **Fluke Ti9 Thermal Imager**

Accuracy: $\pm 5^{\circ}\text{C}$ or 5% of Reading
(whichever is greater)

Measurement Temp Range: -20°C to $+250^{\circ}\text{C}$
(Not calibrated under -10°C)

Date: 6/14/2011

Certified By: **Fluke QA Manager**

Fluke Corporation certifies the above instrumentation has been calibrated to meet or exceed the published specifications. The calibration was performed using instrumentation and standards that are traceable to the United States National Institute of Standards and Technology (NIST), intrinsic standards, or national metrology institutes that actively participates in the activities of the BIPM, directly or indirectly through regional groups. This Certificate/Report shall not be reproduced, except in full, without the written consent of Fluke Corporation.